Technical Letter

An Attempt of pH Mapping on Biological Specimens by pH-Imaging Microscope Using a Semiconductor Silicon Sensor

Yukiko Nakano*, Satoshi Nomura**, Tadafumi Kawamoto***, Junji Tagami****, Atsushi Horiba** and Yoshiro Takano*

* Biostructural Science, Department of Hard Tissue Engineering, Graduate School of Tokyo Medical and Dental University, 1-5-45, Yushima, Bunkyo-ku, Tokyo 113-8549 ** Horiba Ltd., 2, Kisshoinmiyanohigashimachi, Minami-Ku, Kyoto 601-8510 *** RI Research Center, School of Dentistry, Tsurumi University, 2-1-3, Tsurumi-ku, Yokohama 230-0063 **** Cariology and Operative Dentistry, Department of Restorative Sciences, Creducte School of Talva Medical and Dental University, 4, 5, 45, Wushime

Graduate School of Tokyo Medical and Dental University, 1-5-45, Yushima, Bunkyo-ku, Tokyo 113-8549

(Received 18 January 2002, Accepted 2 February 2002)

An attempt of pH mapping of biological samples was made on the whole body sections of fresh frozen rats by means of the pH-imaging microscope using a semiconductor silicon sensor. Fine tuning of experimental conditions for data acquisition and its digital processing successfully yielded pH images at the tissue and organ levels. Present data suggest usefulness of the application of the pH-imaging microscope for the analysis of local pH values of animal tissues that have a potential importance in regulation of biological activities.

Keywards : pH; pH-imaging; frozen section; rat; histrogy.

